WARNING

1. Do not use sling if a patient suffers from back injury.
2. Please understand hoist operating instructions before using.
3. Check sling before use for signs of wear.

Full Body Sling With Divided Leg Lifting Procedures

KH408N1—20kg-40kg
KH408N2—40kg-150kg
KH408N3—150kg-250kg

STEP 1: When the patient is lying in bed, roll them onto their side and place the sling under them similar to fitting a draw sheet, ie: roll the patient onto their side concertina the sling under the patients back. Align (A) with the base of the patients spine, then roll the patient onto the sling and centralise. If the patient is sitting up in bed or in a chair, lean the patient forward and slide the sling down the back of the bed/ chair. Try to tuck (A) as low as possible under the patient bottom.

STEP 2: Bring the lifter into position, ensure hoist legs are divided and the hoist yoke is approximately over the centre of the patient.
DO NOT engage the brakes on the castors

STEP 3: Attach straps (B) & (C) to yoke. The leg pieces (D) & (E) can be attached in three different ways

Attachment 1: Pass (D) & (E) under the thighs and attach to the same side hooks. This causes the legs to be abduct when lifting, this way of attaching the KH408N is useful for toileting but may be uncomfortable for patients without muscle tone.

Attachment 2: Pass (D) & (E) under the thighs and cross over the straps and attach to the opposite side hook. This allows the legs to be supported together but divided.

Attachment 3: Pass (D) & (E) under both thighs and attach to opposite side hooks.

Once (D) & (E) are fixed then attach (F) & (G)

Patients can be positioned 2 ways.

1. Sitting upright—(B) & (C) use short straps, (D) & (E) use long straps, use (F) & (G)
2. Reclined lift—(B) & (C) use long straps, (D) & (E) use short straps, fit (F) & (G)

Note:The operator/career can steady the patient by holding the back handles.

STEP 4. Start with the lifting process using the slow speed. Position the patient over chair, bed or toilet and begin lowering. Make sure the patient feels comfortable and secure in the sling.
CAUTION: Hoist may move towards the patient when boom is raised.

CARE INSTRUCTIONS

1. Sling Care

Protect from direct sunlight.
Do not puncture with any sharp objects.
Repair any tear immediately.
Do not lift a patient if there are any sling defects.
Washing and drying temperature must not exceed 60 degrees.

2. Spot Cleaning

The sling is machine washable.
Treat spills and stains as soon as possible.
Gently scrape away any soils or mop away liquid from surface of fabric with damp cloth.
Do not scrub with stiff brush.
Dry in shade, away from direct heat.
Dry thoroughly before use.

Kerry Equipment™
A Hills Company

4 Halley Rd
Balcatta WA 6021
Tel: 08 9240 2944
Fax: 08 9240 1469
Email: info@kerryequipment.com.au
Website: www.kerryequipment.com.au

PATIENT RECORDS FOR MOUNTING POINTS
USING KH408N1, KH408N2, KH408N3

Name of Patient	Sling size	Laying	Seated	Strap colour (B) & (C)	Strap colour (D) & (E)

USERS GUIDE
KH408N
FULL BODY SLING WITH DIVIDED LEG

FEATURES
Mesh/netting option KH408NE.
Easy to fit.
Full body and head support with divided leg.
Ideal for toileting patients that require full body support.
Padding fitted to leg area.
Commode aperture.
Colour coded attachment straps.
Positioning straps.
Machine washable.

Kerry Equipment™
A Hills Company